



National Laboratory for HIV Reference Services
National HIV and Retrovirology Laboratories
National Microbiology Laboratory
Public Health Agency of Canada

HTLV Serology Quality Assessment Program Summary for Panel HTLVSER 2018Oct26

2018Oct26 HTLV Serology Panel		
Panel Sample	True Status	Labs Reporting Incorrect Status
A	HTLV-II Ab Positive	
B	HTLV-I Ab Positive	
C	Negative	
D	Negative	
E	HTLV-I Ab Positive	

All participants were able to provide either the correct serology status and/or recommendation.



National Laboratory for HIV Reference Services
National HIV and Retrovirology Laboratories
National Microbiology Laboratory
Public Health Agency of Canada

HTLV Serology Quality Assessment Program

Final Report for Panel HTVLSE 2018Oct26

Issued 2019-01-16

Introduction

The NLHRS distributed the 2018Oct26 and 2019Apr16 panels on October 10th, 2018. This final report is specific to the 2018Oct26 panel only and is publicly available; however the identity of participants is not disclosed.

Panel Samples, HTLV Test Kits, and Data Entry

- *Panel Composition*
 - 2018Oct26 HTLV Serology Panel: Five samples; two HTLV negative (C, D), two HTLV-I positive (B, E), and one HTLV-II positive sample (A). Testing and characterization by the NLHRS are presented in Appendix 1. Panels were sent to 18 participants including the NLHRS on October 10th, 2018. The data entry deadline for the 2018Oct26 panel was October 26th, 2018.
- *HTLV Test Kits* – Five different assays were used by the 17 participants excluding the NLHRS (Figure 1). The majority of participants, 88% (15/17), performed screen testing only. One laboratory performed confirmatory testing in the absence of a screen test.
- *Data entry* - The NLHRS Quality Assessment Program (QAP) used the web based Survey Monkey system to capture results. The format of the Final Interpretation section in Survey Monkey was changed to simplify the submission process. Participants were also asked to pilot a new NLHRS QAP website that will replace Survey Monkey in the future.




Figure 1: Assays used by the participants in the NLHRS 2018Oct26 HTLV serology panel (excludes the NLHRS).

Homogeneity and stability

- The homogeneity and stability of the 2018Oct26 HTLV serology panel was assessed by comparing the participants' results (including the NLHRS) with the results of the panel's characterization performed by the NLHRS prior to the test event.
- There is no indication of heterogeneity or instability of the panel samples as the results submitted by the participants are consistent with the expected results from the NLHRS characterization of each panel member (Table 1 and Appendix 1).
- The source material (Access Biological) for the positive panel members is the same source material used for the 2017-2018 HTLV serology panels.

External QC and QA activities

1. *External quality control (QC) material* - Used in addition to controls provided in kits; allows users to detect technical problems and assay sensitivity from lot to lot.
 - Nine participants (53%, 9/17) reported using external QC material.




Figure 2: Source of external quality control used for the 2018Oct26 HTLV serology panel.

2. **Quality Assurance (QA) programs**- Allows participants to evaluate their overall use of the assay and reporting of the results.

- o Thirteen participants (76.4%, 13/17) reported participation in other quality assurance programs (Figure 3).




Figure 3: Distribution of external quality assurance programs which participants are enrolled in other than the NLHRS QAP.

Participants’ feedback collected from Survey Monkey and the beta testing of the new QAP website

- o Of the 17 participants, 16 provided feedback in Survey Monkey. Thirteen participants liked the changes made to the survey compared to the previous iteration (Figure 4).
- o Several areas of improvement for the next survey were identified by the participants (Figure 5).
- o Six participants were satisfied with the current format while 3 participants had no comments regarding areas the NLHRS could improve upon (Figure 5).
- o All participants participated in the beta testing of the new NLHRS QAP website. Feedback on the new NLHRS QAP website is still being collected.
- o Suggestions collected in Survey Monkey will be incorporated into the new NLHRS QAP website which will streamline the results entry process and improve overall functionality.




Figure 4: Number of participants’ who a) liked the changes to survey monkey and b) used the new QAP website




Figure 5: Participants’ responses to which area requires improvement in the NLHRS HTLV serology survey.

Legend: Major Intermediate Minor

Table 1: 2018Oct26 HTLV panel final status reported from participants (includes the NLHRS).					
LAB	SAMPLE A HTLV-II Ab Positive	SAMPLE B HTLV-I Ab Positive	SAMPLE C Negative	SAMPLE D Negative	SAMPLE E HTLV-I Ab Positive
HV01	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV02	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV03	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV12	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV15	HTLV-II Ab Positive	HTLV-I Ab Positive	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I Ab Positive
HV16	HTLV-II Ab Positive	HTLV-I Ab Positive	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I Ab Positive
HV17	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV18	Would not report based on results ¹	Would not report based on results ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	Would not report based on results ¹
HV20	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV21	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV22	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV44	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV50	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV55	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV63	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV75	HTLV-II Ab Positive	HTLV-I Ab Positive	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I Ab Positive
HV76	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹
HV80	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Positive ¹	HTLV-I/II Ab Negative	HTLV-I/II Ab Negative	HTLV-I/II Ab Positive ¹

¹ Further action recommended by participant; “Refer for further HTLV testing or request follow-up samples”.

Table 2: Level of the different flags and the causes of the flag	
Level of flag	Causes for flagging
Major	Incorrect result/status provided
Intermediate	Deviation from kit insert, unresolved status without recommendation
Minor	Minor errors that do not result in misinterpretation of the true status of the sample, unresolved status but made a recommendation

Results (Excluding the NLHRS)

- *Return rate*
 - 100% of the participants returned results by the deadline (17/17).
- *Qualitative Group Analysis (Table 1)*
 - *Sample A (HTLV-II Ab Positive)* – All participants provided either a correct serology status and/or recommendation.
 - *Sample B (HTLV-I Ab Positive)* – All participants provided either a correct serology status and/or recommendation.
 - *Sample C (Negative)* – All participants provided either a correct serology status and/or recommendation.
 - *Sample D (Negative)* – All participants provided either a correct serology status and/or recommendation.
 - *Sample E (HTLV-I Ab Positive)* – All participants provided either a correct serology status and/or recommendation.

Discussion

All participants were able to correctly identify the HTLV-I Ab positive and the HTLV-II Ab positive samples either through an HTLV screening assay or HTLV confirmatory assay. Similarly, samples C and D were correctly identified as HTLV Ab negative by all participants. In addition, no post-analytical errors were made.

The NLHRS made some minor changes to the 2018Oct26 HTLV serology survey such as changing the results selection format in the Final Interpretation section in Survey Monkey. The format was changed to a drop down menu instead of the point and click format used in previous iterations of the survey. This change was made in order to minimize the likelihood of a participant making an incorrect selection based on the results submission format. Although the majority of participants were satisfied with the changes made in Survey Monkey, we recognize that participants would like further improvements to the reporting system. To address this, the NLHRS is in the process of implementing a new results submission website that will resolve issues inherent to the current Survey Monkey submission system.

Conclusion

The absence of any errors found in the 2018Oct26 HTLV Serology panel is not surprising as all participants have consistently demonstrated good technical and post analytical competency throughout each NLHRS HTLV serology proficiency test event.

The NLHRS would like to express our gratitude to those that participated in the beta testing of the new NLHRS QAP website. Your feedback will be used to finalize the submission website before it is fully implemented.

We value each laboratory's participation in these QA panels and your suggestions for improvement. The NLHRS is committed to improving all aspects of the HTLV serology proficiency testing program in order to provide quality proficiency testing to our participants.

The overall quality of HTLV antibody testing in Canada remains very high.

If you have any comments or concerns please contact us at:

phac.nlhrs.qap-peq.lnsrv.aspc@canada.ca

Thank you for your participation in the NLHRS HTLV Serology Quality Assurance Program



John Ho
Quality Assurance Program Coordinator
National Laboratory for HIV Reference Services
Public Health Agency of Canada
Tel: (204) 789-6522



Dr. John Kim
Laboratory Chief
National Laboratory for HIV Reference Services
Public Health Agency of Canada
Tel: (204) 789-6527

Appendix 1: Characterization

Summary of NLHRS Characterization of the 2018Oct26 HTLV Panel Samples

The NLHRS 2018Oct26 HTLV Panel Sample Testing Results									
Sample	Final Status	NLHRS Testing							
		Fujirebio INNO-LIA HTLV I/II Score							
		Interpretation	p19 I/II	p24 I/II	gp46 I/II	gp21 I/II	p19 I	gp46 I	gp46 II
A	HTLV-II Ab Positive	HTLV-II Positive	++	+++	+++	++	-	-	+++
B	HTLV-I Ab Positive	HTLV-I Positive	+++	+++	+++	+++	++	+++	-
C	Negative	Negative	-	-	-	-	-	-	-
D	Negative	Negative	-	-	-	-	-	-	-
E	HTLV-I Ab Positive	HTLV-I Positive	+++	+++	+++	+++	++	+++	-

N/T: Not tested

Appendix 2: Troubleshooting

Troubleshooting; common causes of outlying and/or aberrant results in Serology and Molecular Laboratories.

Type of Error	Possible Cause(s)	Pre-Analytical	Analytical	Post- Analytical
Sample mix-up	Can occur during specimen reception or testing. May result in outlying/aberrant results for one or all samples mixed-up.	✓	✓	
Transcription	• Incorrect test ordering by physician	✓		
	• Incorrect shipment address	✓		
	• Selecting the wrong assay for data entry	✓		
	• Interchanging results for two or more specimens			✓
	• Entering incorrect results			✓
	• Entering values in the incorrect field (e.g., OD as S/Co)			✓
	• Entering values in the incorrect unit (e.g., IU/mL instead of log ₁₀ copies/mL)			✓
	• Using a comma instead of a dot to denote a decimal point			✓
	• Selecting the incorrect assay interpretation or analyte			✓
	• Failure to recommend follow-up testing where necessary			✓
It is recommended all results that are manually transcribed or entered electronically be checked by a second individual to avoid transcription errors.				
Outlying and/or Aberrant Results (random error)	<u>Sporadic test results identified as outlying and/or aberrant can be classified as random events. Possible causes of random error include:</u>			
	• Incorrect sample storage/shipping conditions	✓	✓	
	• Incorrect test method	✓	✓	
	• Insufficient mixing of sample, especially following freezing		✓	
	• Poor pipetting		✓	
	• Ineffective or inconsistent washing		✓	
	• Transcription errors	✓		✓
	• Cross-contamination or carryover	✓	✓	
• Presence of inhibitors to PCR		✓		
Outlying and/or Aberrant Results (systematic error)	<u>A series of test results identified as outlying and/or aberrant may be due to a systematic problem. Systematic problems may be due to:</u>			
	• Reagents contaminated, expired, or subject to batch variation		✓	
	• Instrument error or malfunction		✓	
	• Insufficient washing		✓	
	• Incorrect wavelength used to read the assay result		✓	
	• Cycling times too long/short or temperature too high/low		✓	
	• Incubation time too long/short or temperature too high/low		✓	
	• Insufficient mixing/centrifuging before testing		✓	
	• Incorrect storage of test kits and/or reagents	✓		
	• Contamination of master-mix, extraction areas or equipment		✓	
	• Ineffective extraction process		✓	
	• Degradation of master-mix components		✓	
• Suboptimal primer design (in-house assays)		✓		

This table was modified from a report produced by the National Reference Laboratory (NRL), Melbourne, Australia.